

## INFORMATION REPORT

REPORT

CD NO.

COUNTRY East Germany

SUBJECT Exposure Device for X-Ray Intensity Wedge

DATE DISTR. 15 March 1954

NO. OF PAGES 2

25X1

PLACE  
ACQUIREDNO. OF ENCLS.  
(LISTED BELOW)DATE OF  
INFO.SUPPLEMENT TO  
REPORT NO.

THIS DOCUMENT CONTAINS INFORMATION AFFECTING THE NATIONAL DEFENSE  
OF THE UNITED STATES WITHIN THE MEANING OF TITLE 18, SECTIONS 793  
AND 794, OF THE U. S. CODE, AS AMENDED. ITS TRANSMISSION OR REVEL-  
ATION OF ITS CONTENTS TO OR RECEIPT BY AN UNAUTHORIZED PERSON  
IS PROHIBITED BY LAW. THE REPRODUCTION OF THIS FORM IS PROHIBITED.

THIS IS UNEVALUATED INFORMATION

25X1

1. Research work in the Department for Crystal Structure Analysis of the Academy Institute for Medicine and Biology in Berlin-Buch has been hampered by the lack of an adequate method for determining the intensity of diffraction spots (called "reflexes" in Buch terminology) in the X-ray film of a crystal.

25X1 this shortcoming is a serious handicap for crystal research analysis quite generally and prevails everywhere. So far, determination of the intensity had to be carried out by visual estimate, and this procedure involved a margin of error of from five to thirty percent. While the relatively small error of five percent is attainable at the beginning of an observation, it gradually increases to thirty percent as a result of fatigue caused by long observation time.

25X12. development of a device which is supposed to eliminate to a great extent the error due to subjective estimate. The device is called "The Exposure Device for X-Ray Intensity Wedge" (Belichtungsgeraet fuer Roentgenintensitaets-keil); construction plans for the device were finished in late November 1953. With the aid of this device and an evaluation device,

25X1 expects to reduce the margin of error in the determination of X-ray intensity to plus/minus 0.5 percent.

3. The Exposure Device works in the following way: The device is provided with a dark-slide (Kassette) into which the X-ray film of a crystal is introduced. The dark-slide has two diaphragms which can be closed and opened by means of two slides. The diaphragms are exposed to an X-ray source after the slides are opened. By rotating the diaphragms two black "wedges" are produced on the X-ray film of the crystal. The position of the diaphragms is such that the blackening of the X-ray film only takes place on the margin of the film, so that the X-ray picture of the crystal is not disturbed. The black "wedges" are not uniformly black, but have different shades of blackness. With the aid of the evaluation device yet to be developed it will be possible to compare the intensities of the diffraction spots with the "wedges" and express them in exact figures. The dark-slide is protected by a sheet

|       |   |      |   |      |  |              |  |        |   |
|-------|---|------|---|------|--|--------------|--|--------|---|
| STATE | X | NAVY | X | NSRB |  | DISTRIBUTION |  | OSI Ev | X |
| ARMY  | X | AIR  | X | FBI  |  |              |  |        |   |

-2-

of lead against the X-ray source, so that direct exposure and exposure to scattered rays is impossible. The dimensions of the dark-slide have been selected so that films of several types of X-ray cameras, such as the Weissenberg and De Jong cameras, can be inserted.

- 25X1 4. Construction of the exposure device has begun in the research laboratory  
25X1 of the Institute for Medicine and Biology. [REDACTED]

✓ In addition to the considerably increased  
25X1 accuracy in the determination of intensities, the following advantages  
25X1 will result from the device [REDACTED]

- a. Determination of intensities can be performed by non-expert,  
auxiliary personnel;
- b. Duration of the evaluation of an X-ray picture will be reduced by  
about eighty percent.
5. The complete device, including the evaluation device, is supposed to be  
in operating order by the end of 1954.

25X1

[REDACTED]

-2-

25X1

SECRET

[REDACTED]

[REDACTED]